



# EU DECLARATION OF CONFORMITY

**HEINE Optotechnik GmbH & Co. KG**  
**Dornierstr. 6, 82205 Gilching, Germany**  
www.heine.com

Single Registration Number: DE-MF-000006269

Medical device

**Product family:** Sphygmomanometers  
**Product group:** GAMMA XXL

We hereby declare, under our sole responsibility, the conformity of the following product in accordance with MDD 93/42/EEC.

<b>Product name</b>	GAMMA XXL LF
<b>Basic UDI-DI</b>	4053755_AS_02_48
<b>GMDN</b>	16156
<b>UMDNS</b>	16-156
<b>EMDN</b>	C9006
<b>Classification</b>	Class Im according annex IX



HEINE Optotechnik GmbH & Co. KG hereby declares that the product covered by this declaration is in conformity with this Regulation and, where applicable, with other relevant EU legislation providing for the issuing of an EU declaration of conformity.

References to any common specifications: None


Conformity assessment procedure chosen: MDD 93/42/EEC, annex VII in combination with annex V

This declaration of conformity is valid until a revised declaration of conformity is issued.

According to article 120 section 3 of the MDR (EU) 2017/745, a transition period for the sphygmomanometers is available. The sphygmomanometers could be placed on the market until 31 December 2028 under the directive 93/42/EEC (see annex on page 2 ff.).

**HEINE OPTOTECHNIK**  
GmbH & Co. KG  
Dornierstr. 6  
82205 Gilching

Gilching, 15 January 2024  
(Place and date of issue)

  
Thomas Sauerer / PRRC  
(Name/function and signature)

## HEINE Optotechnik GmbH & Co. KG

Dornierstr. 6  
82205 Gilching  
Germany

Date: 2023.12.21

### Notified Body Confirmation Letter

Reference: 1000156810

To whom it may concern,

**Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices**

This letter confirms that, DQS Medizinprodukte GmbH, a Notified Body designated against Regulation (EU) 2017/745 (MDR) and identified by the number 0297 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

## HEINE Optotechnik GmbH & Co. KG

Dornierstr. 6  
82205 Gilching  
Germany

SRN: DE-MF-000006269

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables listed below: Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which DQS Medizinprodukte GmbH is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but DQS Medizinprodukte GmbH has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry, or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices. The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices

- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body,

i.A. **Hovsep Aro**  
Regulatory Affairs Manager

A handwritten signature in black ink, appearing to be 'Hovsep Aro', written over a horizontal line.

**Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

Device name and Basic UDI-DI (as proposed by the manufacturer within the application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<b>GAMMA G5</b> Basic UDI-DI: 4053755_AS_01_45	Class I devices with a measuring function	N/A	325735 MR5 (NB 0297)
<b>GAMMA G7</b> Basic UDI-DI: 4053755_AS_01_45	Class I devices with a measuring function	N/A	325735 MR5 (NB 0297)
<b>GAMMA GP</b> Basic UDI-DI: 4053755_AS_01_45	Class I devices with a measuring function	N/A	325735 MR5 (NB 0297)
<b>GAMMA XXL LF</b> Basic UDI-DI: 4053755_AS_02_48	Class I devices with a measuring function	N/A	325735 MR5 (NB 0297)

**Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

Device name and Basic UDI-DI (as proposed by the manufacturer within the application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
N/A	N/A	N/A	N/A

#### Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2023-12-21	1000156810	Initial issue
	Cert-ID	description of change (e.g. addition of device XYZ to Table 1)
	Cert-ID	description of change (e.g. removal of device XYZ from Table 2)



# EC-CERTIFICATE

(Production quality assurance)



This is to certify that the company



## HEINE Optotechnik GmbH & Co. KG

Dornierstr. 6  
82205 Gilching  
Germany

has implemented and maintains a quality assurance system which applies to the manufacture and final controls of the products.

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of

### Annex V of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

Sphygmomanometers and Accessories (Class Im)

The manufacturer is subject to surveillance according to Annex V, Section 4. The CE marking with the Notified Body Identification Number (0297) may be affixed on the devices listed in the certificate. The certificate is in the case of class I(s) devices (I(s) = class I products placed on the market in sterile conditions) limited to the aspects of manufacture concerned with securing and maintaining sterile conditions. The certificate is in the case of class I(m) devices (I(m) = class I devices with a measuring function) limited to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

Certificate registration no.	325735 MR5
Certificate unique ID	170770401
Effective date	2020-07-31
Expiry date	2024-01-27
Frankfurt am Main	2020-07-31

### DQS Medizinprodukte GmbH

Sigrid Uhlemann  
Managing Director

Dr. Thomas Feldmann  
Head of Certification Body

August-Schanz-Straße 21, 60433 Frankfurt am Main,  
Tel. +49 (0) 69 95427-300, [medical.devices@dqs-med.de](mailto:medical.devices@dqs-med.de)

DQS Medizinprodukte GmbH is a Notified Body according to Council Directive 93/42/EEC concerning medical devices with the Identification Number 0297.

