

## Reusable versus Single-use Anaesthetic Equipment Life Cycle Cost and Environmental Savings

**Financial and environmental costs of reusable and single-use anaesthetic equipment.**  
McGain F, Story D, Lim T, McAlister S. Br J Anaesth. 2017;118(6):862-869.

Life Cycle Assessment is an internationally, standardized, science based approach to **quantify multiple environmental and public health impacts of a product and/or process over its entire life span**, from raw material extraction, device production, transport, use and reuse, reprocessing and disposal. Life Cycle Assessment should **support procurement in decision making** on the **real costs** and **environmental burden** of a product.

This study aims to assess in two Australian hospital with 6 operating rooms (OR) real costs from changing from reusable to disposable anaesthetic equipment, including various different multi-use vs single-use anaesthetic equipment combinations scenarios\*

### Significant life cycle and environmental savings of re-usable vs. mainly single-use anaesthetic equipment.



#### Environmental Savings:

Greenhouse Gas Emissions (GHG) measured in CO2 equivalents  
Depends on **energy source**, up to 50% (US) p.a. or similar in Australia due to mainly coal sourced energy used

**~85% p.a.**



#### Annual Operating Cost Savings:

All multi-use vs. mainly single-use in 2015 with 6 operating rooms

**~46% p.a.**  
**- 32.033 AU\$**

**“In all five scenarios the financial cost to process single-use anaesthetic equipment was more than for reusable anaesthetic equipment.”**

#### Important quotes:

- > “If all UK hospitals had single use anaesthetics and would convert to reusables this would be the equivalent of **taking >1000 cars off UK roads**”.
- > “Given that **reusable direct laryngoscope blades could be washed with surgical equipment in surgical-type washer loads, the consequences of replacing them with single-use variants would not lead to any measureable reduction in the number of washer loads** required”.
- > “We did **not include washer and sterilizer maintenance and depreciation as these are fixed annual costs**. These costs would be unaltered by the presence or absence of reusable anaesthetic equipment, owing to the much larger number of surgical items that would require cleaning regardless”.
- > “**Labour costs to process all reusable equipment were modest**. Most **environmental impacts to process anaesthetic equipment were small**, with only CO2 emissions and water use being relatively important”.

\*Scenario 1 All Multi-use (MU) anaesthetic circuits/ face masks and Laryngeal mask airways (LMA), direct and videolaryngoscope blades and handles, Scenario 2: Mainly Single Use (SU) anaesthetic circuits / face masks / LMA / Blades, but MU handles & videolaryngoscopes, Scenario 3 All SU (Model), Scenario 4 Mainly MU, only SU. <https://www.greenhospitals.net/wp-content/uploads/2017/09/Comparing-Reusable-to-Single-use-Anesthetic-Equipment-Australia.pdf>, accessed on Oct 18th 2017 face masks (Model), Scenario 5 Mainly MU, only SU blades (Model).